

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041619\
 Data File : VX008996.D
 Acq On : 16 Apr 2019 18:34
 Operator : JC/SP
 Sample : K2440-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW-6

Quant Time: Apr 17 06:02:06 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 16 09:35:21 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	259001	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	410428	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	358573	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	156939	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	165849	45.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.88%	
35) Dibromofluoromethane	5.50	113	121278	46.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.08%	
50) Toluene-d8	8.71	98	514513	48.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.96%	
62) 4-Bromofluorobenzene	11.13	95	176659	47.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.28%	
Target Compounds						
16) Acetone	2.44	43	8625	4.557	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041619\
Data File : VX008996.D
Acq On : 16 Apr 2019 18:34
Operator : JC/SP
Sample : K2440-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW-6

Quant Time: Apr 17 06:02:06 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 16 09:35:21 2019
Response via : Initial Calibration

