

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041619\
 Data File : VX008997.D
 Acq On : 16 Apr 2019 18:57
 Operator : JC/SP
 Sample : K2440-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW-7

Quant Time: Apr 17 06:04:03 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 16 09:35:21 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	265591	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	423534	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	364008	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	154080	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	170467	45.54	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.08%	
35) Dibromofluoromethane	5.50	113	123631	45.48	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.96%	
50) Toluene-d8	8.71	98	522596	48.21	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.42%	
62) 4-Bromofluorobenzene	11.13	95	172979	45.20	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.40%	
Target Compounds						
16) Acetone	2.44	43	10335	5.325	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041619\
Data File : VX008997.D
Acq On : 16 Apr 2019 18:57
Operator : JC/SP
Sample : K2440-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW-7

Quant Time: Apr 17 06:04:03 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 16 09:35:21 2019
Response via : Initial Calibration

