

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041619\
 Data File : VX009001.D
 Acq On : 16 Apr 2019 20:30
 Operator : JC/SP
 Sample : K2411-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-16-S

Quant Time: Apr 17 06:41:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 16 09:35:21 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	265139	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	415442	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	371302	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	161883	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	170198	45.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.10%	
35) Dibromofluoromethane	5.50	113	125313	47.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.00%	
50) Toluene-d8	8.71	98	527691	49.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.26%	
62) 4-Bromofluorobenzene	11.13	95	180654	48.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.26%	
Target Compounds						
16) Acetone	2.44	43	11622	5.998	ug/l	97
20) Methylene Chloride	2.85	84	17512	5.285	ug/l	91
25) 2-Butanone	4.68	43	11235	3.627	ug/l	91
43) Isopropyl Acetate	6.45	43	16238	1.823	ug/l	100
95) Naphthalene	13.83	128	54334	4.391	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041619\
Data File : VX009001.D
Acq On : 16 Apr 2019 20:30
Operator : JC/SP
Sample : K2411-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-16-S

Quant Time: Apr 17 06:41:55 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 16 09:35:21 2019
Response via : Initial Calibration

