

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041620\
 Data File : VX015705.D
 Acq On : 16 Apr 2020 12:37
 Operator : JC/SP
 Sample : L2276-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB180-TB-20200409

Quant Time: Apr 17 07:23:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1053141	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	1610759	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1500002	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	808265	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	624398	42.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.08%	
35) Dibromofluoromethane	5.47	113	484128	47.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.04%	
50) Toluene-d8	8.70	98	1901213	50.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.30%	
62) 4-Bromofluorobenzene	11.12	95	785372	50.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.96%	
Target Compounds						
64) Tetrachloroethene	9.32	164	3781	0.334	ug/l #	74
93) 1,2,4-Trichlorobenzene	13.63	180	5562	0.284	ug/l	98
94) Hexachlorobutadiene	13.76	225	3302	0.354	ug/l	96
96) 1,2,3-Trichlorobenzene	14.01	180	4806	0.255	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041620\
Data File : VX015705.D
Acq On : 16 Apr 2020 12:37
Operator : JC/SP
Sample : L2276-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB180-TB-20200409

Quant Time: Apr 17 07:23:17 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

