

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041620\
 Data File : VX015706.D
 Acq On : 16 Apr 2020 12:59
 Operator : JC/SP
 Sample : L2233-19
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z3-DUP-0543-200408

Quant Time: Apr 17 07:30:07 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1067137	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1592652	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1544365	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	836908	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	629335	41.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.62%	
35) Dibromofluoromethane	5.47	113	490404	48.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.34%	
50) Toluene-d8	8.70	98	1925004	51.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.72%	
62) 4-Bromofluorobenzene	11.12	95	797165	52.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.66%	
Target Compounds						
21) trans-1,2-Dichloroethene	3.15	96	9927	0.888	ug/l	86
27) cis-1,2-Dichloroethene	4.57	96	444789	34.384	ug/l	88
44) Trichloroethene	7.19	130	247096	21.077	ug/l	91
64) Tetrachloroethene	9.32	164	3422519	293.654	ug/l	98
71) Bromoform	10.84	173	8293	0.740	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041620\
Data File : VX015706.D
Acq On : 16 Apr 2020 12:59
Operator : JC/SP
Sample : L2233-19
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z3-DUP-0543-200408

Quant Time: Apr 17 07:30:07 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

