

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041620\
 Data File : VX015710.D
 Acq On : 16 Apr 2020 14:30
 Operator : JC/SP
 Sample : L2293-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-200413

Quant Time: Apr 17 07:54:44 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1078941	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	1605721	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1557608	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	847956	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	629416	41.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.72%	
35) Dibromofluoromethane	5.47	113	487785	47.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.04%	
50) Toluene-d8	8.70	98	1918345	50.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.52%	
62) 4-Bromofluorobenzene	11.12	95	818038	53.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.52%	
Target Compounds						
44) Trichloroethene	7.20	130	5279	0.447	ug/l	95
64) Tetrachloroethene	9.32	164	7410	0.630	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041620\
Data File : VX015710.D
Acq On : 16 Apr 2020 14:30
Operator : JC/SP
Sample : L2293-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
TB-01-200413

Quant Time: Apr 17 07:54:44 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

