

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041620\
 Data File : VX015713.D
 Acq On : 16 Apr 2020 15:38
 Operator : JC/SP
 Sample : L2293-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z5-S1SVEGW-200407

Quant Time: Apr 17 08:08:16 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1066213	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1631246	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1561168	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	873865	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	633970	42.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.32%	
35) Dibromofluoromethane	5.47	113	492750	47.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.52%	
50) Toluene-d8	8.70	98	1931841	50.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.64%	
62) 4-Bromofluorobenzene	11.12	95	825003	52.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.76%	
Target Compounds						
3) Chloromethane	1.31	50	6745	0.564	ug/l	94
16) Acetone	2.42	43	119131	18.440	ug/l	98
25) 2-Butanone	4.65	43	37236	3.584	ug/l	95
30) Chloroform	5.19	83	20256	0.959	ug/l	92
47) Bromodichloromethane	7.88	83	31664	1.881	ug/l	94
59) 2-Hexanone	9.48	43	15521	1.098	ug/l	93
60) Dibromochloromethane	9.56	129	47397	3.547	ug/l	95
64) Tetrachloroethene	9.31	164	4082	0.346	ug/l #	68
71) Bromoform	10.84	173	39242	3.466	ug/l #	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041620\
Data File : VX015713.D
Acq On : 16 Apr 2020 15:38
Operator : JC/SP
Sample : L2293-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z5-S1SVEGW-200407

Quant Time: Apr 17 08:08:16 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

