

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041620\
 Data File : VX015717.D
 Acq On : 16 Apr 2020 17:08
 Operator : JC/SP
 Sample : L2276-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB180-GW-318-320

Quant Time: Apr 17 08:30:08 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1021080	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	1589670	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1525948	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	825013	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	621324	43.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.28%	
35) Dibromofluoromethane	5.47	113	482265	47.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.92%	
50) Toluene-d8	8.70	98	1885178	50.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.78%	
62) 4-Bromofluorobenzene	11.12	95	777026	51.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.22%	
Target Compounds						
16) Acetone	2.42	43	24661	3.986	ug/l	95
17) Carbon Disulfide	2.55	76	10951	0.372	ug/l #	92
52) Toluene	8.77	92	5897	0.223	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041620\
Data File : VX015717.D
Acq On : 16 Apr 2020 17:08
Operator : JC/SP
Sample : L2276-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB180-GW-318-320

Quant Time: Apr 17 08:30:08 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

