

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041620\
 Data File : VX015718.D
 Acq On : 16 Apr 2020 17:31
 Operator : JC/SP
 Sample : L2276-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB180-GW-338-340

Quant Time: Apr 17 08:32:51 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1024067	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	1534226	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1486624	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	806069	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	604314	41.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.68%	
35) Dibromofluoromethane	5.47	113	468498	47.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.54%	
50) Toluene-d8	8.70	98	1844200	51.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.16%	
62) 4-Bromofluorobenzene	11.12	95	760014	51.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.58%	
Target Compounds						
3) Chloromethane	1.31	50	4740	0.413	ug/l	99
16) Acetone	2.42	43	32613	5.256	ug/l	95
17) Carbon Disulfide	2.55	76	18943	0.642	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041620\
Data File : VX015718.D
Acq On : 16 Apr 2020 17:31
Operator : JC/SP
Sample : L2276-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB180-GW-338-340

Quant Time: Apr 17 08:32:51 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

