

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041620\
 Data File : VX015723.D
 Acq On : 16 Apr 2020 19:24
 Operator : JC/SP
 Sample : L2294-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS042MW130-200413

Quant Time: Apr 17 08:44:43 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1059506	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1654046	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1603930	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	899778	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	646227	43.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.50%	
35) Dibromofluoromethane	5.47	113	503248	47.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.20%	
50) Toluene-d8	8.70	98	1981564	50.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.80%	
62) 4-Bromofluorobenzene	11.12	95	826585	52.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.50%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.57	96	374074	29.125	ug/l	88
44) Trichloroethene	7.19	130	204022	16.757	ug/l	97
64) Tetrachloroethene	9.32	164	38955	3.218	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041620\
Data File : VX015723.D
Acq On : 16 Apr 2020 19:24
Operator : JC/SP
Sample : L2294-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS042MW130-200413

Quant Time: Apr 17 08:44:43 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

