

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041620\
 Data File : VX015724.D
 Acq On : 16 Apr 2020 19:46
 Operator : JC/SP
 Sample : L2294-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS042MW032-200413

Quant Time: Apr 17 08:46:33 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1100030	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	6.83	114	1626868	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1575971	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	899315	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	637518	41.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.18%	
35) Dibromofluoromethane	5.47	113	487983	46.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.84%	
50) Toluene-d8	8.70	98	1921261	50.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.36%	
62) 4-Bromofluorobenzene	11.12	95	820872	52.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.50%	
Target Compounds						
16) Acetone	2.42	43	7334	1.100	ug/l	97
27) cis-1,2-Dichloroethene	4.56	96	47484	3.561	ug/l	85
44) Trichloroethene	7.19	130	32049	2.676	ug/l	95
64) Tetrachloroethene	9.32	164	11465	0.964	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041620\
Data File : VX015724.D
Acq On : 16 Apr 2020 19:46
Operator : JC/SP
Sample : L2294-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS042MW032-200413

Quant Time: Apr 17 08:46:33 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

