

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041620\
 Data File : VX015725.D
 Acq On : 16 Apr 2020 20:09
 Operator : JC/SP
 Sample : L2294-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z3-DUP-0545-200413

Quant Time: Apr 17 08:49:23 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	1115269	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1669661	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1588431	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	907424	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	644040	40.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.88%	
35) Dibromofluoromethane	5.47	113	507431	47.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.08%	
50) Toluene-d8	8.70	98	2001192	50.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.86%	
62) 4-Bromofluorobenzene	11.12	95	830093	51.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.96%	
Target Compounds						
16) Acetone	2.42	43	12210	1.807	ug/l	96
27) cis-1,2-Dichloroethene	4.57	96	4403	0.326	ug/l	87
64) Tetrachloroethene	9.32	164	4627	0.386	ug/l #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041620\
Data File : VX015725.D
Acq On : 16 Apr 2020 20:09
Operator : JC/SP
Sample : L2294-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z3-DUP-0545-200413

Quant Time: Apr 17 08:49:23 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

