

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041624\
 Data File : VX041097.D
 Acq On : 16 Apr 2024 11:43
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Apr 17 01:42:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041624W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 01:32:05 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	180193	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	266648	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	261623	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	149137	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	95244	48.957	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.920%
35) Dibromofluoromethane	5.385	113	85868	48.206	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.420%
50) Toluene-d8	8.647	98	314287	50.373	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.740%
62) 4-Bromofluorobenzene	11.079	95	119046	50.263	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.520%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041624\
Data File : VX041097.D
Acq On : 16 Apr 2024 11:43
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Apr 17 01:42:02 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041624W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 01:32:05 2024
Response via : Initial Calibration

