

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041624\
 Data File : VX041105.D
 Acq On : 16 Apr 2024 15:19
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Apr 17 02:11:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041624W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 02:04:44 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	176703	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	264448	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	262224	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	152590	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	98842	51.810	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 103.620%	
35) Dibromofluoromethane	5.385	113	86128	48.755	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 97.500%	
50) Toluene-d8	8.647	98	309825	50.071	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.140%	
62) 4-Bromofluorobenzene	11.079	95	120373	51.246	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 102.500%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041624\
Data File : VX041105.D
Acq On : 16 Apr 2024 15:19
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Apr 17 02:11:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041624W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 02:04:44 2024
Response via : Initial Calibration

