

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041624\
 Data File : VX041112.D
 Acq On : 16 Apr 2024 18:01
 Operator : JC/MD
 Sample : P2113-04 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 403

Quant Time: Apr 17 02:13:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041624W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 02:04:44 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	169912	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	258057	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	257537	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	146394	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	96760	52.746	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 105.500%	
35) Dibromofluoromethane	5.385	113	84323	48.915	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 97.840%	
50) Toluene-d8	8.647	98	306700	50.793	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 101.580%	
62) 4-Bromofluorobenzene	11.079	95	115933	50.578	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 101.160%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041624\
Data File : VX041112.D
Acq On : 16 Apr 2024 18:01
Operator : JC/MD
Sample : P2113-04 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
403

Quant Time: Apr 17 02:13:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041624W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 02:04:44 2024
Response via : Initial Calibration

