

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041624\
 Data File : VX041107.D
 Acq On : 16 Apr 2024 16:06
 Operator : JC/MD
 Sample : P2115-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-5

Quant Time: Apr 17 02:12:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041624W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 02:04:44 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	169850	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	254006	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	243587	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	132210	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	95162	51.894	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 103.780%	
35) Dibromofluoromethane	5.385	113	82732	48.758	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 97.520%	
50) Toluene-d8	8.647	98	296476	49.883	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.760%	
62) 4-Bromofluorobenzene	11.079	95	111100	49.242	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 98.480%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041624\
Data File : VX041107.D
Acq On : 16 Apr 2024 16:06
Operator : JC/MD
Sample : P2115-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-5

Quant Time: Apr 17 02:12:21 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041624W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 02:04:44 2024
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 719
Delta R.T. -0.000 min
Lab File: VX041107.D
Acq: 16 Apr 2024 16:06




Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-5


Tgt Ion:168 Resp: 169850
Ion Ratio Lower Upper
168 100
99 46.5 56.6 85.0#



#33
1,2-Dichloroethane-d4
Concen: 51.894 ug/l
RT: 5.958 min Scan# 799
Delta R.T. 0.006 min
Lab File: VX041107.D
Acq: 16 Apr 2024 16:06

Tgt Ion: 65 Resp: 95162
Ion Ratio Lower Upper
65 100
67 52.0 0.0 100.2









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.763 min Scan# 911
Delta R.T. -0.000 min
Lab File: VX041107.D
Acq: 16 Apr 2024 16:06




Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-5


Tgt Ion:114 Resp: 254006
Ion Ratio Lower Upper
114 100
63 18.8 0.0 45.0
88 15.3 0.0 36.6



#35
Dibromofluoromethane
Concen: 48.758 ug/l
RT: 5.385 min Scan# 705
Delta R.T. -0.000 min
Lab File: VX041107.D
Acq: 16 Apr 2024 16:06

Tgt Ion:113 Resp: 82732
Ion Ratio Lower Upper
113 100
111 102.3 82.9 124.3
192 25.7 13.3 19.9#









#50
Toluene-d8
Concen: 49.883 ug/l
RT: 8.647 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX041107.D
Acq: 16 Apr 2024 16:06




Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-5


Tgt Ion: 98 Resp: 296476
Ion Ratio Lower Upper
98 100
100 66.2 53.1 79.7



#62
4-Bromofluorobenzene
Concen: 49.242 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX041107.D
Acq: 16 Apr 2024 16:06

Tgt Ion: 95 Resp: 111100
Ion Ratio Lower Upper
95 100
174 98.2 0.0 135.6
176 96.1 0.0 131.0





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.055 min Scan# 1471

Delta R.T. -0.000 min

Lab File: VX041107.D

Acq: 16 Apr 2024 16:06

Instrument :

MSVOA_X

ClientSampleId :

STORAGE-BLANK-SOIL-REF-5





Tgt Ion:117 Resp: 243587

Ion Ratio Lower Upper

117 100

82 49.1 46.2 69.4

119 32.2 25.5 38.3



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.024 min Scan# 1794

Delta R.T. -0.000 min

Lab File: VX041107.D

Acq: 16 Apr 2024 16:06

Tgt Ion:152 Resp: 132210

Ion Ratio Lower Upper

152 100

115 53.3 43.3 129.9

150 157.6 0.0 345.0

