

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041719\
 Data File : VX009011.D
 Acq On : 17 Apr 2019 13:50
 Operator : JC/SP
 Sample : K2387-03DL 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS040MW318-190410DL

Quant Time: Apr 18 02:06:22 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:47:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	264451	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	416327	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	368945	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	160228	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	171555	46.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.06%	
35) Dibromofluoromethane	5.50	113	125318	46.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.80%	
50) Toluene-d8	8.71	98	523747	49.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.30%	
62) 4-Bromofluorobenzene	11.13	95	177547	47.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.40%	
Target Compounds						
4) Vinyl Chloride	1.40	62	6893	1.894	ug/l	99
27) cis-1,2-Dichloroethene	4.60	96	56482	16.446	ug/l	90
44) Trichloroethene	7.21	130	27329	8.781	ug/l	97
64) Tetrachloroethene	9.34	164	86963	31.027	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041719\
Data File : VX009011.D
Acq On : 17 Apr 2019 13:50
Operator : JC/SP
Sample : K2387-03DL 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS040MW318-190410DL

Quant Time: Apr 18 02:06:22 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:47:58 2019
Response via : Initial Calibration

