

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041719\
 Data File : VX009018.D
 Acq On : 17 Apr 2019 16:34
 Operator : JC/SP
 Sample : K2448-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS037TW143-190415

Quant Time: Apr 18 03:18:35 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:47:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	264003	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	423237	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	367054	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	153276	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	171589	46.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.24%	
35) Dibromofluoromethane	5.50	113	126899	46.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.44%	
50) Toluene-d8	8.71	98	523022	48.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.56%	
62) 4-Bromofluorobenzene	11.13	95	174230	45.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.12%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	19335	5.639	ug/l	88
44) Trichloroethene	7.21	130	2435	0.770	ug/l	96
64) Tetrachloroethene	9.34	164	3878	1.391	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041719\
Data File : VX009018.D
Acq On : 17 Apr 2019 16:34
Operator : JC/SP
Sample : K2448-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS037TW143-190415

Quant Time: Apr 18 03:18:35 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:47:58 2019
Response via : Initial Calibration

