

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041719\
 Data File : VX009023.D
 Acq On : 17 Apr 2019 18:30
 Operator : JC/SP
 Sample : K2449-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS036MW009-190415

Quant Time: Apr 18 03:34:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:47:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	256752	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	411657	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	351388	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	143684	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	167565	46.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.62%	
35) Dibromofluoromethane	5.50	113	122038	46.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.38%	
50) Toluene-d8	8.71	98	510386	48.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.88%	
62) 4-Bromofluorobenzene	11.13	95	167212	44.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.92%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	3345	1.003	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041719\
Data File : VX009023.D
Acq On : 17 Apr 2019 18:30
Operator : JC/SP
Sample : K2449-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS036MW009-190415

Quant Time: Apr 18 03:34:30 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:47:58 2019
Response via : Initial Calibration

