

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041719\  
 Data File : VX009024.D  
 Acq On : 17 Apr 2019 18:54  
 Operator : JC/SP  
 Sample : K2449-02  
 Misc : 5.0mL/MSVOA X/WATER  
 ALS Vial : 22 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 SS042MW032-190415

Quant Time: Apr 18 03:45:48 2019  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X041519W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Apr 17 06:47:58 2019  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units  | Dev(Min) |
|-----------------------------|-------|------|----------|--------|--------|----------|
| 1) Pentafluorobenzene       | 5.67  | 168  | 260218   | 50.00  | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.86  | 114  | 416887   | 50.00  | ug/l   | 0.00     |
| 63) Chlorobenzene-d5        | 10.11 | 117  | 362402   | 50.00  | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.08 | 152  | 152275   | 50.00  | ug/l   | 0.00     |
| System Monitoring Compounds |       |      |          |        |        |          |
| 33) 1,2-Dichloroethane-d4   | 6.06  | 65   | 169750   | 46.29  | ug/l   | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 92.58% |          |
| 35) Dibromofluoromethane    | 5.50  | 113  | 124566   | 46.56  | ug/l   | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 93.12% |          |
| 50) Toluene-d8              | 8.71  | 98   | 514884   | 48.26  | ug/l   | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 96.52% |          |
| 62) 4-Bromofluorobenzene    | 11.13 | 95   | 174616   | 46.36  | ug/l   | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 92.72% |          |
| Target Compounds            |       |      |          |        |        |          |
| 27) cis-1,2-Dichloroethene  | 4.60  | 96   | 54471    | 16.119 | ug/l   | 92       |
| 44) Trichloroethene         | 7.22  | 130  | 2106     | 0.676  | ug/l   | 90       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041719\  
Data File : VX009024.D  
Acq On : 17 Apr 2019 18:54  
Operator : JC/SP  
Sample : K2449-02  
Misc : 5.0mL/MSVOA X/WATER  
ALS Vial : 22 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
SS042MW032-190415

Quant Time: Apr 18 03:45:48 2019  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_X\METHOD\82X041519W.M  
Quant Title : SW846 8260  
QLast Update : Wed Apr 17 06:47:58 2019  
Response via : Initial Calibration

