

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041719\
 Data File : VX009026.D
 Acq On : 17 Apr 2019 19:41
 Operator : JC/SP
 Sample : K2449-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY037MW144-190415

Quant Time: Apr 18 04:10:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:47:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	248566	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	400154	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	348410	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	146597	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	163486	46.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.34%	
35) Dibromofluoromethane	5.50	113	119392	46.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.98%	
50) Toluene-d8	8.71	98	496210	48.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.90%	
62) 4-Bromofluorobenzene	11.13	95	166256	45.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.98%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	4339	1.344	ug/l	87
44) Trichloroethene	7.21	130	8288	2.771	ug/l	100
64) Tetrachloroethene	9.34	164	8291	3.132	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041719\
Data File : VX009026.D
Acq On : 17 Apr 2019 19:41
Operator : JC/SP
Sample : K2449-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY037MW144-190415

Quant Time: Apr 18 04:10:11 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:47:58 2019
Response via : Initial Calibration

