

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041719\
 Data File : VX009028.D
 Acq On : 17 Apr 2019 20:27
 Operator : JC/SP
 Sample : K2449-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-0510-190415

Quant Time: Apr 18 04:15:40 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:47:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	250000	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	399771	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	347806	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	148496	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	163080	46.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.58%	
35) Dibromofluoromethane	5.50	113	119368	46.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.06%	
50) Toluene-d8	8.71	98	498996	48.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.54%	
62) 4-Bromofluorobenzene	11.13	95	165240	45.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.50%	
Target Compounds						
64) Tetrachloroethene	9.34	164	9847	3.727	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041719\
Data File : VX009028.D
Acq On : 17 Apr 2019 20:27
Operator : JC/SP
Sample : K2449-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-0510-190415

Quant Time: Apr 18 04:15:40 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:47:58 2019
Response via : Initial Calibration

