

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041719\
 Data File : VX009017.D
 Acq On : 17 Apr 2019 16:10
 Operator : JC/SP
 Sample : K2450-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OUTFALL001-190416

Quant Time: Apr 18 10:54:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:47:58 2019
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	259099	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	416332	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	358378	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	151764	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	168751	46.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.42%	
35) Dibromofluoromethane	5.50	113	122609	45.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.78%	
50) Toluene-d8	8.71	98	514696	48.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.60%	
62) 4-Bromofluorobenzene	11.13	95	173008	45.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.98%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.59	96	3226	0.959	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041719\
Data File : VX009017.D
Acq On : 17 Apr 2019 16:10
Operator : JC/SP
Sample : K2450-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OUTFALL001-190416

Quant Time: Apr 18 10:54:48 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:47:58 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.67 min Scan# 740

Delta R.T. -0.00 min

Lab File: VX009017.D

Acq: 17 Apr 2019 16:10

Instrument :

MSVOA_X

ClientSampleId :


OUTFALL001-190416

Tot Ion:168 Resp: 259099

Ion Ratio Lower Upper

168 100

99 54.6 43.0 64.6



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.67

100000

80000



60000

40000

20000

0

Time--> 5.50 5.60 5.70 5.80



#27

cis-1,2-Dichloroethene

Concen: 0.959 ug/l

RT: 4.59 min Scan# 564

Delta R.T. -0.01 min

Lab File: VX009017.D

Acq: 17 Apr 2019 16:10


Tgt Ion: 96 Resp: 3226

Ion Ratio Lower Upper

96 100

61 147.9 0.0 322.0

98 65.7 0.0 128.4



Abundance Ion 95.90 (95.60 to 96.60): VX0

Ion 60.90 (60.60 to 61.60): VX0

Ion 97.90 (97.60 to 98.60): VX0

2000


1500


1000

500

0

Time--> 4.50 4.55 4.60 4.65 4.70









#33
 1,2-Dichloroethane-d4
 Concen: 46.215 ug/l
 RT: 6.06 min Scan# 805
 Delta R.T. -0.00 min
 Lab File: VX009017.D
 Acq: 17 Apr 2019 16:10




Instrument :
 MSVOA_X
 ClientSampleId :
 OUTFALL001-190416


Tot Ion: 65 Resp: 168751
 Ion Ratio Lower Upper
 65 100
 67 54.3 0.0 108.6



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.86 min Scan# 936
 Delta R.T. -0.00 min
 Lab File: VX009017.D
 Acq: 17 Apr 2019 16:10

Tgt Ion: 114 Resp: 416332
 Ion Ratio Lower Upper
 114 100
 63 22.9 0.0 46.0
 88 16.3 0.0 32.6





#35

Dibromofluoromethane

Concen: 45.888 ug/l

RT: 5.50 min Scan# 713

Delta R.T. -0.00 min

Lab File: VX009017.D

Acq: 17 Apr 2019 16:10

Instrument :

MSVOA_X

ClientSampleId :

OUTFALL001-190416


Tot Ion:113 Resp: 122609

Ion Ratio Lower Upper

113 100

111 103.6 82.9 124.3




192 21.6 17.4 26.0



Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V



#50

Toluene-d8

Concen: 48.302 ug/l

RT: 8.71 min Scan# 1240

Delta R.T. -0.00 min

Lab File: VX009017.D


Acq: 17 Apr 2019 16:10

Tgt Ion: 98 Resp: 514696

Ion Ratio Lower Upper



98 100


100 64.0 51.4 77.2



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): VX0





#62

4-Bromofluorobenzene

Concen: 45.994 ug/l

RT: 11.13 min Scan# 1637

Delta R.T. -0.00 min

Lab File: VX009017.D

Acq: 17 Apr 2019 16:10

Instrument :

MSVOA_X

ClientSampleId :

OUTFALL001-190416


Tot Ion: 95 Resp: 173008

Ion Ratio Lower Upper

95 100

174 78.3 0.0 159.6




176 76.2 0.0 154.8



Abundance Ion 94.90 (94.60 to 95.60): VX008953.D (-1630) (-)

Ion 173.80 (173.50 to 174.50): VX008953.D (-1630) (-)

Ion 175.80 (175.50 to 176.50): VX008953.D (-1630) (-)



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.12 min Scan# 1470

Delta R.T. 0.01 min

Lab File: VX009017.D

Acq: 17 Apr 2019 16:10


Tgt Ion: 117 Resp: 358378

Ion Ratio Lower Upper

117 100

82 57.5 48.2 72.4



119 31.8 25.0 37.6




Abundance Ion 116.90 (116.60 to 117.60): VX008953.D (-1462) (-)

Ion 82.00 (81.70 to 82.70): VX008953.D (-1462) (-)

Ion 118.90 (118.60 to 119.60): VX008953.D (-1462) (-)





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.08 min Scan# 1792

Delta R.T. -0.00 min

Lab File: VX009017.D

Acq: 17 Apr 2019 16:10

Instrument :

MSVOA_X

ClientSampleId :

OUTFALL001-190416


Tot Ion:152 Resp: 151764

Ion Ratio Lower Upper

152 100

115 59.0 41.4 124.2

150 157.3 0.0 349.8



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

