

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX041720\
 Data File : VX015752.D
 Acq On : 17 Apr 2020 20:33
 Operator : JC/SP
 Sample : L2305-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Apr 20 01:20:22 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	986413	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	1539059	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1518077	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	844959	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.04	65	584010	41.98	ug/l	0.00
Spiked Amount			Recovery	=	83.96%	
35) Dibromofluoromethane	5.47	113	469439	47.72	ug/l	0.00
Spiked Amount			Recovery	=	95.44%	
50) Toluene-d8	8.70	98	1856043	51.24	ug/l	0.00
Spiked Amount			Recovery	=	102.48%	
62) 4-Bromofluorobenzene	11.12	95	799169	54.29	ug/l	0.00
Spiked Amount			Recovery	=	108.58%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Chloromethane	1.56	50	892105	80.608	ug/l #	44
4) Vinyl Chloride	1.18	62	28693	2.824	ug/l #	47
5) Bromomethane	1.57	94	3156	Below Cal	#	10
6) Chloroethane	1.51	64	81839	12.985	ug/l #	51
10) Methyl Iodide	2.50	142	3854	0.322	ug/l #	86
11) Tert butyl alcohol	3.01	59	29301	8.799	ug/l #	69
13) Acrolein	2.25	56	9612	5.067	ug/l #	40
14) Allyl chloride	2.68	41	38771	1.782	ug/l #	46
16) Acetone	2.42	43	3725333	623.274	ug/l	96
17) Carbon Disulfide	2.55	76	22266637	783.781	ug/l #	85
18) Methyl Acetate	2.75	43	10384	0.681	ug/l #	69
25) 2-Butanone	4.64	43	288348	29.998	ug/l	100
29) Tetrahydrofuran	5.10	42	14144	2.289	ug/l #	73
30) Chloroform	5.18	83	17984	0.921	ug/l	94
36) 1,1-Dichloropropene	5.63	75	67545	4.763	ug/l #	51
37) Ethyl Acetate	4.64	43	288348	15.617	ug/l #	71
38) Carbon Tetrachloride	5.64	117	87547	5.700	ug/l #	14
40) Benzene	6.12	78	166718	4.014	ug/l	100
43) Isopropyl Acetate	6.31	43	21726	0.697	ug/l #	66
49) 1,4-Dioxane	7.70	88	148	0.510	ug/l #	20
51) 4-Methyl-2-Pentanone	8.62	43	5644	0.320	ug/l #	56
52) Toluene	8.76	92	12038	0.470	ug/l	96
55) 1,1,2-Trichloroethane	9.07	97	13671	1.303	ug/l #	18
59) 2-Hexanone	9.48	43	3116	0.234	ug/l	81
68) m/p-Xylenes	10.34	106	8837	0.442	ug/l	90
69) o-Xylene	10.68	106	4502	0.228	ug/l	95
76) 1,2,3-Trichloropropane	11.12	75	390472	22.472	ug/l #	35
81) trans-1,4-Dichloro-2-buten	11.12	75	390472	53.921	ug/l #	10
86) p-Isopropyltoluene	12.05	119	12238	0.247	ug/l #	73
95) Naphthalene	13.82	128	283660	4.544	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041720\
Data File : VX015752.D
Acq On : 17 Apr 2020 20:33
Operator : JC/SP
Sample : L2305-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Apr 20 01:20:22 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.63 min Scan# 778

Delta R.T. -0.00 min

Lab File: VX015752.D

Acq: 17 Apr 2020 20:33

Instrument :

MSVOA_X


ClientSampleId :

Tot Ion:168 Resp: 986413

Ion Ratio Lower Upper




168 100

99 53.0 46.4 69.6



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#3

Chloromethane

Concen: 80.608 ug/l

RT: 1.56 min Scan# 111

Delta R.T. 0.25 min

Lab File: VX015752.D


Acq: 17 Apr 2020 20:33

Tgt Ion: 50 Resp: 892105

Ion Ratio Lower Upper



50 100


52 0.3 25.3 37.9#



Abundance Ion 49.90 (49.60 to 50.60): VX0

Ion 51.90 (51.60 to 52.60): VX0





#4

Vinyl Chloride

Concen: 2.824 ug/l

RT: 1.18 min Scan# 48

Delta R.T. -0.21 min

Lab File: VX015752.D

Acq: 17 Apr 2020 20:33





Instrument :
MSVOA_X
ClientSampled :

Tot Ion: 62 Resp: 28693

Ion Ratio Lower Upper

62 100

64 2.8 26.0 39.0#



#5

Bromomethane

Concen: Below Cal

RT: 1.57 min Scan# 112

Delta R.T. -0.05 min

Lab File: VX015752.D




Acq: 17 Apr 2020 20:33


Tgt Ion: 94 Resp: 3156

Ion Ratio Lower Upper

94 100

96 7.2 74.8 112.2#





#6

Chloroethane

Concen: 12.985 ug/l

RT: 1.51 min Scan# 102

Delta R.T. -0.20 min

Lab File: VX015752.D

Acq: 17 Apr 2020 20:33

Instrument :

MSVOA_X





ClientSampleId :

Tot Ion: 64 Resp: 81839

Ion Ratio Lower Upper

64 100

66 4.6 25.4 38.2#



#10

Methyl Iodide

Concen: 0.322 ug/l

RT: 2.50 min Scan# 264

Delta R.T. 0.01 min

Lab File: VX015752.D

Acq: 17 Apr 2020 20:33




Tgt Ion: 142 Resp: 3854


Ion Ratio Lower Upper

142 100

127 41.0 35.9 53.9

141 30.2 11.6 17.4#









#11

Tert butyl alcohol
Concen: 8.799 ug/l
RT: 3.01 min Scan# 349
Delta R.T. -0.00 min
Lab File: VX015752.D
Acq: 17 Apr 2020 20:33

Instrument :
MSVOA_X
ClientSampleId :




Tot Ion: 59 Resp: 29301
Ion Ratio Lower Upper
59 100
57 21.7 8.1 12.1#




#13

Acrolein
Concen: 5.067 ug/l
RT: 2.25 min Scan# 224
Delta R.T. -0.02 min
Lab File: VX015752.D
Acq: 17 Apr 2020 20:33

Tgt Ion: 56 Resp: 9612
Ion Ratio Lower Upper
56 100
55 21.6 57.0 85.4#





#14

Allvl chloride

Concen: 1.782 ug/l

RT: 2.68 min Scan# 294

Delta R.T. -0.02 min

Lab File: VX015752.D

Acq: 17 Apr 2020 20:33

Instrument :
MSVOA_X
ClientSampleId :


Tot Ion: 41 Resp: 38771

Ion Ratio Lower Upper




41 100

39 21.2 51.4 77.0#

76 0.0 21.8 32.8#



Abundance Ion 41.00 (40.70 to 41.70): VX0
Ion 39.00 (38.70 to 39.70): VX0
Ion 75.90 (75.60 to 76.60): VX0



#16

Acetone

Concen: 623.274 ug/l

RT: 2.42 min Scan# 251

Delta R.T. -0.00 min

Lab File: VX015752.D


Acq: 17 Apr 2020 20:33

Tgt Ion: 43 Resp: 3725333



Ion Ratio Lower Upper


43 100

58 29.5 22.1 33.1



Abundance Ion 43.00 (42.70 to 43.70): VX0
Ion 58.00 (57.70 to 58.70): VX0





#17

Carbon Disulfide

Concen: 783.781 ug/l

RT: 2.55 min Scan# 272

Delta R.T. -0.01 min

Lab File: VX015752.D

Acq: 17 Apr 2020 20:33

Instrument :

MSVOA_X





ClientSampleId :

Tot Ion: 76 Resp: 22266637

Ion Ratio Lower Upper

76 100

78 14.2 7.1 10.7#



#18

Methyl Acetate

Concen: 0.681 ug/l

RT: 2.75 min Scan# 305

Delta R.T. -0.01 min

Lab File: VX015752.D




Acq: 17 Apr 2020 20:33


Tgt Ion: 43 Resp: 10384

Ion Ratio Lower Upper

43 100

74 35.1 16.5 24.7#





#25

2-Butanone

Concen: 29.998 ug/l

RT: 4.64 min Scan# 615

Delta R.T. -0.00 min

Lab File: VX015752.D

Acq: 17 Apr 2020 20:33





Instrument :
MSVOA_X
ClientSampled :

Tot Ion: 43 Resp: 288348

Ion Ratio Lower Upper

43 100

72 23.8 19.0 28.4



#29

Tetrahydrofuran

Concen: 2.289 ug/l

RT: 5.10 min Scan# 691

Delta R.T. 0.01 min

Lab File: VX015752.D

Acq: 17 Apr 2020 20:33




Tgt Ion: 42 Resp: 14144


Ion Ratio Lower Upper

42 100

72 21.9 31.7 47.5#

71 22.1 29.6 44.4#









#30
Chloroform
Concen: 0.921 ug/l
RT: 5.18 min Scan# 704
Delta R.T. -0.01 min
Lab File: VX015752.D
Acq: 17 Apr 2020 20:33




Instrument :
MSVOA_X
ClientSampleId :


Tot Ion: 83 Resp: 17984
Ion Ratio Lower Upper
83 100
85 57.4 49.5 74.3



#33
1,2-Dichloroethane-d4
Concen: 41.978 ug/l
RT: 6.04 min Scan# 845
Delta R.T. 0.01 min
Lab File: VX015752.D
Acq: 17 Apr 2020 20:33

Tgt Ion: 65 Resp: 584010
Ion Ratio Lower Upper
65 100
67 52.6 0.0 102.6





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.83 min Scan# 975

Delta R.T. -0.01 min

Lab File: VX015752.D

Acq: 17 Apr 2020 20:33

Instrument :

MSVOA_X

ClientSampleId :


Tot Ion:114 Resp: 1539059

Ion Ratio Lower Upper

114 100

63 21.2 0.0 44.2




88 16.3 0.0 33.8



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 47.716 ug/l

RT: 5.47 min Scan# 752

Delta R.T. -0.00 min

Lab File: VX015752.D

Acq: 17 Apr 2020 20:33


Tgt Ion:113 Resp: 469439

Ion Ratio Lower Upper

113 100

111 104.0 81.5 122.3



192 24.2 17.8 26.6




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V









#36
 1,1-Dichloropropene
 Concen: 4.763 ug/l
 RT: 5.63 min Scan# 778
 Delta R.T. -0.14 min
 Lab File: VX015752.D
 Acq: 17 Apr 2020 20:33




Instrument :
 MSVOA_X
 ClientSampleId :


Tot Ion	Ratio	Lower	Upper
75	100		
110	9.9	17.1	51.3#
77	0.0	24.6	36.8#



#37
 Ethyl Acetate
 Concen: 15.617 ug/l
 RT: 4.64 min Scan# 615
 Delta R.T. -0.16 min
 Lab File: VX015752.D
 Acq: 17 Apr 2020 20:33

Tgt Ion	Ratio	Lower	Upper
43	100		
61	0.4	9.9	14.9#
70	0.2	8.1	12.1#





#38

Carbon Tetrachloride

Concen: 5.700 ug/l

RT: 5.64 min Scan# 779

Delta R.T. -0.12 min

Lab File: VX015752.D





Acq: 17 Apr 2020 20:33

Instrument :

MSVOA_X

ClientSampleId :

Tot Ion:	117	Resp:	87547
Ion	Ratio	Lower	Upper
117	100		
119	3.5	78.0	117.0#
121	0.0	23.8	35.8#



#40

Benzene

Concen: 4.014 ug/l




RT: 6.12 min Scan# 858


Delta R.T. -0.00 min

Lab File: VX015752.D

Acq: 17 Apr 2020 20:33

Tgt Ion:	78	Resp:	166718
Ion	Ratio	Lower	Upper
78	100		
77	24.7	19.8	29.8





#43

Isopropyl Acetate

Concen: 0.697 ug/l

RT: 6.31 min Scan# 889

Delta R.T. -0.11 min

Lab File: VX015752.D

Acq: 17 Apr 2020 20:33

Instrument :

MSVOA_X

ClientSampled :





Tot Ion: 43 Resp: 21726

Ion Ratio Lower Upper

43 100

61 1.4 14.8 22.2#

87 1.0 9.4 14.0#



#49

1,4-Dioxane

Concen: 0.510 ug/l

RT: 7.70 min Scan# 1118

Delta R.T. -0.01 min

Lab File: VX015752.D

Acq: 17 Apr 2020 20:33




Tgt Ion: 88 Resp: 148


Ion Ratio Lower Upper

88 100

43 0.0 30.1 45.1#

58 0.0 60.6 90.8#









#50
Toluene-d8
Concen: 51.242 ug/l
RT: 8.70 min Scan# 1281
Delta R.T. -0.00 min
Lab File: VX015752.D
Acq: 17 Apr 2020 20:33




Instrument :
MSVOA_X
ClientSampled :


Tot Ion: 98 Resp: 1856043
Ion Ratio Lower Upper
98 100
100 67.1 53.5 80.3



#51
4-Methyl-2-Pentanone
Concen: 0.320 ug/l
RT: 8.62 min Scan# 1269
Delta R.T. -0.00 min
Lab File: VX015752.D
Acq: 17 Apr 2020 20:33

Tgt Ion: 43 Resp: 5644
Ion Ratio Lower Upper
43 100
58 63.4 29.7 44.5#





#52

Toluene

Concen: 0.470 ug/l

RT: 8.76 min Scan# 1292

Delta R.T. -0.01 min

Lab File: VX015752.D

Acq: 17 Apr 2020 20:33





Instrument :
MSVOA_X
ClientSampled :

Tot Ion: 92 Resp: 12038

Ion Ratio Lower Upper

92 100

91 176.4 136.6 205.0



#55

1,1,2-Trichloroethane

Concen: 1.303 ug/l

RT: 9.07 min Scan# 1342

Delta R.T. -0.13 min

Lab File: VX015752.D

Acq: 17 Apr 2020 20:33

Tgt Ion: 97 Resp: 13671




Ion Ratio Lower Upper


97 100

83 0.0 70.2 105.4#

85 0.0 44.7 67.1#

99 9.6 51.1 76.7#









#59
2-Hexanone
Concen: 0.234 ug/l
RT: 9.48 min Scan# 1409
Delta R.T. -0.00 min
Lab File: VX015752.D
Acq: 17 Apr 2020 20:33




Instrument :
MSVOA_X
ClientSampled :


Tot Ion: 43 Resp: 3116
Ion Ratio Lower Upper
43 100
58 64.6 25.6 76.6



#62
4-Bromofluorobenzene
Concen: 54.290 ug/l
RT: 11.12 min Scan# 1679
Delta R.T. -0.00 min
Lab File: VX015752.D
Acq: 17 Apr 2020 20:33

Tgt Ion: 95 Resp: 799169
Ion Ratio Lower Upper
95 100
174 83.3 0.0 158.6
176 82.5 0.0 159.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.10 min Scan# 1511
Delta R.T. -0.00 min
Lab File: VX015752.D
Acq: 17 Apr 2020 20:33




Instrument :
MSVOA_X
ClientSampled :


Tot Ion:117 Resp: 1518077
Ion Ratio Lower Upper
117 100
82 54.7 47.8 71.8
119 31.9 26.8 40.2



#68
m/p-Xylenes
Concen: 0.442 ug/l
RT: 10.34 min Scan# 1551
Delta R.T. -0.00 min
Lab File: VX015752.D
Acq: 17 Apr 2020 20:33

Tgt Ion:106 Resp: 8837
Ion Ratio Lower Upper
106 100
91 188.6 163.4 245.0









#69
o-Xylene
Concen: 0.228 ug/l
RT: 10.68 min Scan# 1607
Delta R.T. -0.00 min
Lab File: VX015752.D
Acq: 17 Apr 2020 20:33




Instrument :
MSVOA_X
ClientSampleId :


Tot Ion:106 Resp: 4502
Ion Ratio Lower Upper
106 100
91 222.6 107.1 321.3



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.06 min Scan# 1833
Delta R.T. -0.00 min
Lab File: VX015752.D
Acq: 17 Apr 2020 20:33

Tgt Ion:152 Resp: 844959
Ion Ratio Lower Upper
152 100
115 57.7 39.0 116.9
150 152.9 0.0 340.8









#76
 1,2,3-Trichloropropane
 Concen: 22.472 ug/l
 RT: 11.12 min Scan# 1679
 Delta R.T. -0.16 min
 Lab File: VX015752.D
 Acq: 17 Apr 2020 20:33




Instrument :
 MSVOA_X
 ClientSampleId :


Tot Ion: 75 Resp: 390472
 Ion Ratio Lower Upper
 75 100
 77 1.3 21.2 63.6#



#81
 trans-1,4-Dichloro-2-butene
 Concen: 53.921 ug/l
 RT: 11.12 min Scan# 1679
 Delta R.T. 0.06 min
 Lab File: VX015752.D
 Acq: 17 Apr 2020 20:33

Tgt Ion: 75 Resp: 390472
 Ion Ratio Lower Upper
 75 100
 53 0.2 80.1 120.1#
 89 0.2 36.1 54.1#









#86
p-Isopropyltoluene
Concen: 0.247 ug/l
RT: 12.05 min Scan# 1831
Delta R.T. -0.00 min
Lab File: VX015752.D
Acq: 17 Apr 2020 20:33

Instrument :
MSVOA_X
ClientSampled :

Tot Ion	Ratio	Lower	Upper
119	100		
134	35.8	13.1	39.3
91	43.2	12.4	37.4



#95
Naphthalene
Concen: 4.544 ug/l
RT: 13.82 min Scan# 2121
Delta R.T. -0.00 min
Lab File: VX015752.D
Acq: 17 Apr 2020 20:33

Tgt Ion	Ratio	Lower	Upper
128	100		
127	12.7	10.6	15.8
129	10.3	8.8	13.2

