

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041720\
 Data File : VX015734.D
 Acq On : 17 Apr 2020 13:46
 Operator : JC/SP
 Sample : VX0417MBL01
 Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0417MBL01

Quant Time: Apr 20 03:20:55 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	1035796	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1584267	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1525577	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	840707	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	626315	42.87	ug/l	0.00
Spiked Amount	50.000		Recovery	=	85.74%	
35) Dibromofluoromethane	5.47	113	485436	47.93	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.86%	
50) Toluene-d8	8.70	98	1913623	51.32	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.64%	
62) 4-Bromofluorobenzene	11.12	95	797666	52.64	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.28%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041720\
Data File : VX015734.D
Acq On : 17 Apr 2020 13:46
Operator : JC/SP
Sample : VX0417MBL01
Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0417MBL01

Quant Time: Apr 20 03:20:55 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

