

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041720\
 Data File : VX015739.D
 Acq On : 17 Apr 2020 15:39
 Operator : JC/SP
 Sample : L2281-02ME 10X
 Misc : 4.67g/5mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-38ME

Quant Time: Apr 20 08:46:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1010129	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	1608464	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1525523	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	854415	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	622957	43.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.46%	
35) Dibromofluoromethane	5.47	113	475404	46.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.48%	
50) Toluene-d8	8.70	98	1913249	50.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.08%	
62) 4-Bromofluorobenzene	11.12	95	811125	52.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.44%	
Target Compounds						
68) m/p-Xylenes	10.34	106	70737	3.521	ug/l	99
69) o-Xylene	10.68	106	39383	1.984	ug/l	98
80) 1,3,5-Trimethylbenzene	11.49	105	270110	5.650	ug/l	99
84) 1,2,4-Trimethylbenzene	11.79	105	502111	10.306	ug/l	98
95) Naphthalene	13.82	128	81001	1.283	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041720\
Data File : VX015739.D
Acq On : 17 Apr 2020 15:39
Operator : JC/SP
Sample : L2281-02ME 10X
Misc : 4.67µ/5mL/100µL/5.0mL/MSVOA_X/MEOH
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-38ME

Quant Time: Apr 20 08:46:19 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

