

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041720\
 Data File : VX015744.D
 Acq On : 17 Apr 2020 17:32
 Operator : JC/SP
 Sample : L2298-02DL 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GROUNDWATERDL

Quant Time: Apr 20 05:20:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	991420	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1543180	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1471321	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	786009	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	597123	42.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.40%	
35) Dibromofluoromethane	5.47	113	464386	47.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.16%	
50) Toluene-d8	8.70	98	1819434	50.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.20%	
62) 4-Bromofluorobenzene	11.12	95	752664	50.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.98%	
Target Compounds						
16) Acetone	2.42	43	3235946	538.662	ug/l	97
25) 2-Butanone	4.65	43	39434	4.082	ug/l	94
27) cis-1,2-Dichloroethene	4.57	96	7782	0.648	ug/l	66
95) Naphthalene	13.82	128	68217	1.175	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041720\
Data File : VX015744.D
Acq On : 17 Apr 2020 17:32
Operator : JC/SP
Sample : L2298-02DL 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GROUNDWATERDL

Quant Time: Apr 20 05:20:17 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

