

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041720\
 Data File : VX015745.D
 Acq On : 17 Apr 2020 17:54
 Operator : JC/SP
 Sample : L2293-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z5-S1SVEGW-200407

Quant Time: Apr 20 09:02:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	991009	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1549778	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1481360	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	826754	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	597111	42.72	ug/l	0.00
Spiked Amount	50.000		Recovery	=	85.44%	
35) Dibromofluoromethane	5.47	113	467344	47.17	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.34%	
50) Toluene-d8	8.70	98	1850587	50.74	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.48%	
62) 4-Bromofluorobenzene	11.12	95	778436	52.52	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.04%	
Target Compounds						
16) Acetone	2.42	43	95312	15.872	ug/l	99
25) 2-Butanone	4.65	43	44249	4.582	ug/l #	84
30) Chloroform	5.19	83	20579	1.049	ug/l	100
47) Bromodichloromethane	7.89	83	32687	2.044	ug/l #	95
60) Dibromochloromethane	9.57	129	46453	3.659	ug/l	96
71) Bromoform	10.84	173	36208	3.370	ug/l #	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041720\
Data File : VX015745.D
Acq On : 17 Apr 2020 17:54
Operator : JC/SP
Sample : L2293-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z5-S1SVEGW-200407

Quant Time: Apr 20 09:02:14 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

