

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041720\
 Data File : VX015734.D
 Acq On : 17 Apr 2020 13:46
 Operator : JC/SP
 Sample : VX0417MBL01
 Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0417MBL01

Quant Time: Apr 20 03:20:55 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	1035796	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1584267	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1525577	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	840707	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	626315	42.87	ug/l	0.00
Spiked Amount	50.000		Recovery	=	85.74%	
35) Dibromofluoromethane	5.47	113	485436	47.93	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.86%	
50) Toluene-d8	8.70	98	1913623	51.32	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.64%	
62) 4-Bromofluorobenzene	11.12	95	797666	52.64	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.28%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041720\
Data File : VX015734.D
Acq On : 17 Apr 2020 13:46
Operator : JC/SP
Sample : VX0417MBL01
Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0417MBL01

Quant Time: Apr 20 03:20:55 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.63 min Scan# 778

Delta R.T. 0.00 min

Lab File: VX015734.D

Acq: 17 Apr 2020 13:46

Instrument :

MSVOA_X

ClientSampleId :


VX0417MBL01

Tot Ion:168 Resp: 1035796

Ion Ratio Lower Upper

168 100

99 52.8 46.4 69.6



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.63

400000

300000



200000

100000

0

Time-->

5.50 5.60 5.70 5.80



#33

1,2-Dichloroethane-d4

Concen: 42.873 ug/l

RT: 6.03 min Scan# 844

Delta R.T. 0.00 min

Lab File: VX015734.D


Acq: 17 Apr 2020 13:46

Tgt Ion: 65 Resp: 626315

Ion Ratio Lower Upper

65 100

67 51.9 0.0 102.6



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0

6.03

250000

200000

150000


100000


50000

0

Time-->

5.90 6.00 6.10 6.20





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.84 min Scan# 976

Delta R.T. 0.00 min

Lab File: VX015734.D

Acq: 17 Apr 2020 13:46

Instrument :

MSVOA_X

ClientSampleId :

VX0417MBL01


Tot Ion:114 Resp: 1584267

Ion Ratio Lower Upper

114 100

63 20.3 0.0 44.2




88 15.8 0.0 33.8



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 47.935 ug/l

RT: 5.47 min Scan# 752

Delta R.T. 0.00 min

Lab File: VX015734.D

Acq: 17 Apr 2020 13:46


Tgt Ion:113 Resp: 485436

Ion Ratio Lower Upper

113 100

111 101.5 81.5 122.3



192 23.7 17.8 26.6




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 51.324 ug/l

RT: 8.70 min Scan# 1281

Delta R.T. 0.00 min

Lab File: VX015734.D

Acq: 17 Apr 2020 13:46

Instrument :

MSVOA_X

ClientSampled :





VX0417MBL01

Tot Ion: 98 Resp: 1913623

Ion Ratio Lower Upper

98 100

100 66.3 53.5 80.3



#62

4-Bromofluorobenzene

Concen: 52.642 ug/l

RT: 11.12 min Scan# 1679

Delta R.T. 0.00 min

Lab File: VX015734.D

Acq: 17 Apr 2020 13:46




Tgt Ion: 95 Resp: 797666


Ion Ratio Lower Upper

95 100

174 87.2 0.0 158.6

176 83.0 0.0 159.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.10 min Scan# 1511
Delta R.T. 0.00 min
Lab File: VX015734.D
Acq: 17 Apr 2020 13:46

Instrument :
MSVOA_X
ClientSampled :
VX0417MBL01

Tot Ion:117 Resp: 1525577
Ion Ratio Lower Upper
117 100
82 54.6 47.8 71.8
119 32.0 26.8 40.2



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.06 min Scan# 1833
Delta R.T. 0.00 min
Lab File: VX015734.D
Acq: 17 Apr 2020 13:46

Tgt Ion:152 Resp: 840707
Ion Ratio Lower Upper
152 100
115 55.5 39.0 116.9
150 151.9 0.0 340.8

