

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041723\
 Data File : VX035107.D
 Acq On : 17 Apr 2023 14:29
 Operator : JC/MD
 Sample : 02353-03 200X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BUR-1261

Quant Time: Apr 18 05:53:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 01 01:50:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	310127	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	519002	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	474734	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	220473	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	196744	41.481	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	82.960%
35) Dibromofluoromethane	5.385	113	157950	46.012	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.020%
50) Toluene-d8	8.647	98	616529	48.658	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.320%
62) 4-Bromofluorobenzene	11.079	95	235284	47.986	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.980%
Target Compounds						
					Qvalue	
20) Methylene Chloride	2.788	84	2849	0.749	ug/l	92
52) Toluene	8.720	92	2033	0.214	ug/l	82
94) Hexachlorobutadiene	13.731	225	943	0.519	ug/l	99
95) Naphthalene	13.780	128	3450	0.235	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041723\
Data File : VX035107.D
Acq On : 17 Apr 2023 14:29
Operator : JC/MD
Sample : 02353-03 200X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BUR-1261

Quant Time: Apr 18 05:53:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Quant Title : SW846 8260
QLast Update : Sat Apr 01 01:50:18 2023
Response via : Initial Calibration

