

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041723\
 Data File : VX035110.D
 Acq On : 17 Apr 2023 15:39
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Apr 18 05:53:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 01 01:50:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	302418	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	503773	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	444423	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	187310	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	190881	41.271	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	82.540%
35) Dibromofluoromethane	5.385	113	153815	46.162	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.320%
50) Toluene-d8	8.647	98	593909	48.290	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.580%
62) 4-Bromofluorobenzene	11.079	95	212834	44.719	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.440%
Target Compounds						
20) Methylene Chloride	2.794	84	3813	1.028	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041723\
Data File : VX035110.D
Acq On : 17 Apr 2023 15:39
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Apr 18 05:53:57 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Quant Title : SW846 8260
QLast Update : Sat Apr 01 01:50:18 2023
Response via : Initial Calibration

