

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041725\
 Data File : VX045823.D
 Acq On : 17 Apr 2025 13:12
 Operator : JC/MD
 Sample : Q1818-07 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 3651

Quant Time: Apr 18 03:37:59 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	59360	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	116171	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	107299	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	45186	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	59010	54.360	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.720%	
35) Dibromofluoromethane	5.385	113	41496	50.345	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.680%	
50) Toluene-d8	8.647	98	147797	51.374	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.740%	
62) 4-Bromofluorobenzene	11.079	95	54593	52.097	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 104.200%	
Target Compounds						
					Qvalue	
52) Toluene	8.726	92	1859	0.904	ug/l	95
67) Ethyl Benzene	10.195	91	1494	0.364	ug/l	93
68) m/p-Xylenes	10.305	106	2188	1.465	ug/l	98
69) o-Xylene	10.640	106	638	0.434	ug/l	68
80) 1,3,5-Trimethylbenzene	11.451	105	931	0.311	ug/l	77
84) 1,2,4-Trimethylbenzene	11.756	105	2645	0.880	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041725\
Data File : VX045823.D
Acq On : 17 Apr 2025 13:12
Operator : JC/MD
Sample : Q1818-07 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
3651

Quant Time: Apr 18 03:37:59 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

