

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041725\
 Data File : VX045829.D
 Acq On : 17 Apr 2025 15:32
 Operator : JC/MD
 Sample : Q1810-02 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MOO-25-0118

Quant Time: Apr 18 03:40:14 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	70667	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	135507	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	126596	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	55690	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	67951	52.581	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.160%
35) Dibromofluoromethane	5.379	113	48060	49.988	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.980%
50) Toluene-d8	8.647	98	171494	51.105	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.200%
62) 4-Bromofluorobenzene	11.079	95	66342	54.275	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	108.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041725\
Data File : VX045829.D
Acq On : 17 Apr 2025 15:32
Operator : JC/MD
Sample : Q1810-02 100X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MOO-25-0118

Quant Time: Apr 18 03:40:14 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

