

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041725\
 Data File : VX045830.D
 Acq On : 17 Apr 2025 15:55
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Apr 18 03:40:31 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	78423	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	152261	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	138926	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	59915	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	73546	51.282	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 102.560%	
35) Dibromofluoromethane	5.385	113	53485	49.510	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.020%	
50) Toluene-d8	8.647	98	190259	50.458	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.920%	
62) 4-Bromofluorobenzene	11.079	95	71840	52.306	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 104.620%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041725\
Data File : VX045830.D
Acq On : 17 Apr 2025 15:55
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Apr 18 03:40:31 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

