

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041725\
 Data File : VX045832.D
 Acq On : 17 Apr 2025 16:42
 Operator : JC/MD
 Sample : Q1818-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 3651

Quant Time: Apr 18 03:41:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	70744	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	140247	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	129445	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	55581	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	68297	52.791	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.580%
35) Dibromofluoromethane	5.379	113	48860	49.103	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.200%
50) Toluene-d8	8.647	98	173852	50.056	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.120%
62) 4-Bromofluorobenzene	11.079	95	66564	52.616	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	105.240%
Target Compounds						
						Qvalue
11) Tert butyl alcohol	2.971	59	12926	74.345	ug/l	99
16) Acetone	2.380	43	8440	15.887	ug/l	99
25) 2-Butanone	4.605	43	1133	1.457	ug/l	94
51) 4-Methyl-2-Pentanone	8.580	43	1634	0.961	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041725\
Data File : VX045832.D
Acq On : 17 Apr 2025 16:42
Operator : JC/MD
Sample : Q1818-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
3651

Quant Time: Apr 18 03:41:16 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

