

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX041725\  
 Data File : VX045833.D  
 Acq On : 17 Apr 2025 17:05  
 Operator : JC/MD  
 Sample : Q1814-01  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 18 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 A3729

Quant Time: Apr 18 03:41:39 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X040225W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Apr 02 03:11:43 2025  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units | Dev(Min)  |
|-----------------------------|--------|-------|----------|----------|-------|-----------|
| -----                       |        |       |          |          |       |           |
| Internal Standards          |        |       |          |          |       |           |
| 1) Pentafluorobenzene       | 5.544  | 168   | 74085    | 50.000   | ug/l  | 0.00      |
| 34) 1,4-Difluorobenzene     | 6.757  | 114   | 142315   | 50.000   | ug/l  | 0.00      |
| 63) Chlorobenzene-d5        | 10.049 | 117   | 128735   | 50.000   | ug/l  | 0.00      |
| 72) 1,4-Dichlorobenzene-d4  | 12.018 | 152   | 53112    | 50.000   | ug/l  | 0.00      |
| System Monitoring Compounds |        |       |          |          |       |           |
| 33) 1,2-Dichloroethane-d4   | 5.952  | 65    | 69861    | 51.565   | ug/l  | 0.00      |
| Spiked Amount               | 50.000 | Range | 74 - 125 | Recovery | =     | 103.120%  |
| 35) Dibromofluoromethane    | 5.379  | 113   | 50317    | 49.832   | ug/l  | 0.00      |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =     | 99.660%   |
| 50) Toluene-d8              | 8.647  | 98    | 175704   | 49.854   | ug/l  | 0.00      |
| Spiked Amount               | 50.000 | Range | 86 - 113 | Recovery | =     | 99.700%   |
| 62) 4-Bromofluorobenzene    | 11.079 | 95    | 66454    | 51.766   | ug/l  | 0.00      |
| Spiked Amount               | 50.000 | Range | 77 - 121 | Recovery | =     | 103.540%  |
| Target Compounds            |        |       |          |          |       |           |
| 16) Acetone                 | 2.380  | 43    | 1813     | 3.259    | ug/l  | Qvalue 96 |
| -----                       |        |       |          |          |       |           |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX041725\  
Data File : VX045833.D  
Acq On : 17 Apr 2025 17:05  
Operator : JC/MD  
Sample : Q1814-01  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 18 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
A3729

Quant Time: Apr 18 03:41:39 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X040225W.M  
Quant Title : SW846 8260  
QLast Update : Wed Apr 02 03:11:43 2025  
Response via : Initial Calibration

