

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041725\
 Data File : VX045835.D
 Acq On : 17 Apr 2025 17:49
 Operator : JC/MD
 Sample : Q1818-05 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 3828

Quant Time: Apr 18 03:42:23 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	39688	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	58306	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	31029	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	7323	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	9102	12.541	ug/l	0.01
Spiked Amount	50.000	Range	74 - 125	Recovery	=	25.080%#
35) Dibromofluoromethane	5.397	113	11773	28.459	ug/l	0.02
Spiked Amount	50.000	Range	75 - 124	Recovery	=	56.920%#
50) Toluene-d8	8.647	98	66246	45.880	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.760%
62) 4-Bromofluorobenzene	11.079	95	10681	20.308	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	40.620%#
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	1272	4.268	ug/l	93
30) Chloroform	5.105	83	353	0.341	ug/l #	76
86) p-Isopropyltoluene	12.006	119	4107	8.446	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041725\
Data File : VX045835.D
Acq On : 17 Apr 2025 17:49
Operator : JC/MD
Sample : Q1818-05 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
3828

Quant Time: Apr 18 03:42:23 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

