

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041725\
 Data File : VX045836.D
 Acq On : 17 Apr 2025 18:10
 Operator : JC/MD
 Sample : Q1818-06 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 3829

Quant Time: Apr 18 03:42:46 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	44056	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	66478	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	35454	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	8726	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	10974	13.621	ug/l	0.02
Spiked Amount	50.000	Range	74 - 125	Recovery	=	27.240%#
35) Dibromofluoromethane	5.397	113	14012	29.708	ug/l	0.02
Spiked Amount	50.000	Range	75 - 124	Recovery	=	59.420%#
50) Toluene-d8	8.647	98	73327	44.541	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.080%
62) 4-Bromofluorobenzene	11.085	95	13156	21.939	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	43.880%#
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	1872	5.658	ug/l	97
86) p-Isopropyltoluene	12.006	119	1657	2.860	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041725\
Data File : VX045836.D
Acq On : 17 Apr 2025 18:10
Operator : JC/MD
Sample : Q1818-06 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
3829

Quant Time: Apr 18 03:42:46 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

