



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041725\
Data File : VX045816.D
Acq On : 17 Apr 2025 09:46
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Title : SW846 8260
Last Update : Wed Apr 02 03:11:43 2025



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	23.3	6515	PASS
75	95	30	60	59.6	16638	PASS
95	95	100	100	100.0	27925	PASS
96	95	5	9	6.4	1785	PASS
173	174	0.00	2	0.7	130	PASS
174	95	50	100	67.4	18831	PASS
175	174	5	9	7.5	1404	PASS
176	174	95	101	96.8	18227	PASS
177	176	5	9	6.8	1240	PASS