

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041819\
 Data File : VX009032.D
 Acq On : 18 Apr 2019 10:29
 Operator : JC/SP
 Sample : VX0418MBL01
 Misc : 5.00µ/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0418MBL01

Quant Time: Apr 19 04:37:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:47:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	236937	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	379997	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	313951	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	128217	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	157908	47.29	ug/l	0.00
Spiked Amount						
			Recovery	=	94.58%	
35) Dibromofluoromethane	5.49	113	115419	47.33	ug/l	0.00
Spiked Amount						
			Recovery	=	94.66%	
50) Toluene-d8	8.71	98	464677	47.78	ug/l	0.00
Spiked Amount						
			Recovery	=	95.56%	
62) 4-Bromofluorobenzene	11.13	95	149248	43.47	ug/l	0.00
Spiked Amount						
			Recovery	=	86.94%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041819\
Data File : VX009032.D
Acq On : 18 Apr 2019 10:29
Operator : JC/SP
Sample : VX0418MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0418MBL01

Quant Time: Apr 19 04:37:12 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:47:58 2019
Response via : Initial Calibration

