

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041819\
 Data File : VX009037.D
 Acq On : 18 Apr 2019 12:37
 Operator : JC/SP
 Sample : K2387-07DL 20X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY063MW003-190410DL

Quant Time: Apr 19 04:55:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:47:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	250305	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	408923	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	348013	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	148304	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	168353	47.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.46%	
35) Dibromofluoromethane	5.50	113	122960	46.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.70%	
50) Toluene-d8	8.71	98	502071	47.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.94%	
62) 4-Bromofluorobenzene	11.13	95	165553	44.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.62%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	27830	8.561	ug/l	90
44) Trichloroethene	7.21	130	26280	8.597	ug/l	98
64) Tetrachloroethene	9.34	164	86853	32.852	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041819\
Data File : VX009037.D
Acq On : 18 Apr 2019 12:37
Operator : JC/SP
Sample : K2387-07DL 20X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY063MW003-190410DL

Quant Time: Apr 19 04:55:12 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:47:58 2019
Response via : Initial Calibration

