

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041819\
 Data File : VX009038.D
 Acq On : 18 Apr 2019 13:00
 Operator : JC/SP
 Sample : K2448-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ST006MW015-190415

Quant Time: Apr 19 04:58:15 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:47:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	247361	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	398629	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	345642	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	146055	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	164019	47.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.10%	
35) Dibromofluoromethane	5.50	113	119378	46.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.32%	
50) Toluene-d8	8.71	98	495550	48.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.14%	
62) 4-Bromofluorobenzene	11.13	95	165987	46.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.18%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.40	62	1726	0.507	ug/l	94
16) Acetone	2.44	43	3276	1.812	ug/l #	83
21) trans-1,2-Dichloroethene	3.17	96	1551	0.548	ug/l #	77
27) cis-1,2-Dichloroethene	4.60	96	91643	28.528	ug/l	89
44) Trichloroethene	7.21	130	6926	2.324	ug/l	100
64) Tetrachloroethene	9.34	164	6790	2.586	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041819\
Data File : VX009038.D
Acq On : 18 Apr 2019 13:00
Operator : JC/SP
Sample : K2448-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ST006MW015-190415

Quant Time: Apr 19 04:58:15 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:47:58 2019
Response via : Initial Calibration

