

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041819\
 Data File : VX009042.D
 Acq On : 18 Apr 2019 14:34
 Operator : JC/SP
 Sample : K2449-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS040MW015-190415

Quant Time: Apr 19 09:20:41 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:47:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	235562	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	380616	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	326939	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	137867	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	156163	47.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.08%	
35) Dibromofluoromethane	5.50	113	113538	46.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.96%	
50) Toluene-d8	8.71	98	476937	48.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.92%	
62) 4-Bromofluorobenzene	11.13	95	157639	45.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.68%	
Target Compounds						
11) Tert butyl alcohol	3.04	59	3469	4.708	ug/l	95
83) tert-Butylbenzene	11.77	119	16225	1.927	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041819\
Data File : VX009042.D
Acq On : 18 Apr 2019 14:34
Operator : JC/SP
Sample : K2449-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS040MW015-190415

Quant Time: Apr 19 09:20:41 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:47:58 2019
Response via : Initial Calibration

