

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041819\
 Data File : VX009043.D
 Acq On : 18 Apr 2019 14:58
 Operator : JC/SP
 Sample : K2449-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS040MW303-190415

Quant Time: Apr 19 09:21:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:47:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	238633	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	387157	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	328087	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	135477	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	161924	48.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.30%	
35) Dibromofluoromethane	5.50	113	117650	47.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.70%	
50) Toluene-d8	8.71	98	478559	48.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.60%	
62) 4-Bromofluorobenzene	11.13	95	155277	44.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.78%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.59	96	6269	2.023	ug/l	92
44) Trichloroethene	7.21	130	6944	2.399	ug/l	93
64) Tetrachloroethene	9.34	164	6375	2.558	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041819\
Data File : VX009043.D
Acq On : 18 Apr 2019 14:58
Operator : JC/SP
Sample : K2449-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS040MW303-190415

Quant Time: Apr 19 09:21:27 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:47:58 2019
Response via : Initial Calibration

