

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041819\
 Data File : VX009045.D
 Acq On : 18 Apr 2019 15:45
 Operator : JC/SP
 Sample : K2449-11
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ST008MW146-190416

Quant Time: Apr 19 09:22:40 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:47:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	243020	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	394270	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	340655	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	143200	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	162475	47.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.88%	
35) Dibromofluoromethane	5.50	113	118137	46.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.38%	
50) Toluene-d8	8.71	98	488193	48.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.76%	
62) 4-Bromofluorobenzene	11.13	95	162426	45.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.20%	
Target Compounds						
4) Vinyl Chloride	1.40	62	9164	2.740	ug/l	98
12) 1,1-Dichloroethene	2.37	96	1105	0.424	ug/l	81
21) trans-1,2-Dichloroethene	3.17	96	3237	1.165	ug/l	99
27) cis-1,2-Dichloroethene	4.60	96	550736	174.504	ug/l	90
44) Trichloroethene	7.21	130	182799	62.021	ug/l	98
64) Tetrachloroethene	9.34	164	127142	49.130	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041819\
Data File : VX009045.D
Acq On : 18 Apr 2019 15:45
Operator : JC/SP
Sample : K2449-11
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ST008MW146-190416

Quant Time: Apr 19 09:22:40 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:47:58 2019
Response via : Initial Calibration

