

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041819\
 Data File : VX009048.D
 Acq On : 18 Apr 2019 16:55
 Operator : JC/SP
 Sample : K2449-14
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS037MW304-190416

Quant Time: Apr 19 09:27:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:47:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	235241	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	383949	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	327123	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	137775	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	160123	48.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.60%	
35) Dibromofluoromethane	5.50	113	117520	47.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.38%	
50) Toluene-d8	8.71	98	475012	48.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.68%	
62) 4-Bromofluorobenzene	11.13	95	154527	44.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.10%	
Target Compounds						
21) trans-1,2-Dichloroethene	3.17	96	1356	0.504	ug/l #	78
27) cis-1,2-Dichloroethene	4.60	96	38147	12.487	ug/l	91
44) Trichloroethene	7.21	130	32029	11.159	ug/l	98
64) Tetrachloroethene	9.34	164	371228	149.382	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041819\
Data File : VX009048.D
Acq On : 18 Apr 2019 16:55
Operator : JC/SP
Sample : K2449-14
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS037MW304-190416

Quant Time: Apr 19 09:27:11 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:47:58 2019
Response via : Initial Calibration

