

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041819\
 Data File : VX009049.D
 Acq On : 18 Apr 2019 17:18
 Operator : JC/SP
 Sample : K2449-15
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS037MW301-190416

Quant Time: Apr 19 09:28:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:47:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	235343	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	382369	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	324100	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	134870	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	159595	48.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.24%	
35) Dibromofluoromethane	5.50	113	116568	47.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.00%	
50) Toluene-d8	8.71	98	468787	47.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.80%	
62) 4-Bromofluorobenzene	11.13	95	152738	44.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.42%	
Target Compounds						
4) Vinyl Chloride	1.41	62	188087	58.069	ug/l	99
12) 1,1-Dichloroethene	2.37	96	4852	1.923	ug/l	90
21) trans-1,2-Dichloroethene	3.17	96	2564	0.952	ug/l	99
27) cis-1,2-Dichloroethene	4.60	96	1744765	570.872	ug/l	91
44) Trichloroethene	7.21	130	5875	2.055	ug/l	96
64) Tetrachloroethene	9.34	164	6501	2.640	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041819\
Data File : VX009049.D
Acq On : 18 Apr 2019 17:18
Operator : JC/SP
Sample : K2449-15
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS037MW301-190416

Quant Time: Apr 19 09:28:00 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:47:58 2019
Response via : Initial Calibration

