

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041819\
 Data File : VX009051.D
 Acq On : 18 Apr 2019 18:05
 Operator : JC/SP
 Sample : K2461-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PZ-01

Quant Time: Apr 19 09:29:47 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:47:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	232687	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	380555	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	319051	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	134498	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	157726	48.10	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.20%	
35) Dibromofluoromethane	5.50	113	115166	47.15	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.30%	
50) Toluene-d8	8.71	98	465654	47.81	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.62%	
62) 4-Bromofluorobenzene	11.13	95	149636	43.52	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.04%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041819\
Data File : VX009051.D
Acq On : 18 Apr 2019 18:05
Operator : JC/SP
Sample : K2461-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PZ-01

Quant Time: Apr 19 09:29:47 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:47:58 2019
Response via : Initial Calibration

