

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041819\
 Data File : VX009052.D
 Acq On : 18 Apr 2019 18:29
 Operator : JC/SP
 Sample : K2461-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PZ-02

Quant Time: Apr 19 09:30:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:47:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	229877	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	371443	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	319173	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	135244	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	156415	48.28	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.56%	
35) Dibromofluoromethane	5.50	113	113868	47.77	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.54%	
50) Toluene-d8	8.71	98	461975	48.59	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.18%	
62) 4-Bromofluorobenzene	11.13	95	152324	45.39	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.78%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041819\
Data File : VX009052.D
Acq On : 18 Apr 2019 18:29
Operator : JC/SP
Sample : K2461-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PZ-02

Quant Time: Apr 19 09:30:12 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:47:58 2019
Response via : Initial Calibration

