

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041819\
 Data File : VX009054.D
 Acq On : 18 Apr 2019 19:16
 Operator : JC/SP
 Sample : K2461-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GMW-5

Quant Time: Apr 19 05:41:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:47:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	236723	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	385028	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	333918	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	140200	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	159132	47.70	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.40%	
35) Dibromofluoromethane	5.50	113	115543	46.76	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.52%	
50) Toluene-d8	8.71	98	477321	48.44	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.88%	
62) 4-Bromofluorobenzene	11.13	95	158143	45.46	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.92%	
Target Compounds						
16) Acetone	2.45	43	23099	13.352	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041819\
 Data File : VX009054.D
 Acq On : 18 Apr 2019 19:16
 Operator : JC/SP
 Sample : K2461-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampled :
 GMW-5

Quant Time: Apr 19 05:41:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:47:58 2019
 Response via : Initial Calibration

